



Design Optimization and Integrated Simulation Analysis of a Cable-Driven Ankle Rehabilitation Robot

Norsyakina Husain^{1,2}, Muhammad Nazrin Shah^{1,2}, Khairul Salleh Basaruddin^{1,3}, Muhammad Farzik Ijaz^{4,5,*}, Hiroshi Takemura⁶ and Shafriza Nisha Basah^{1,2}

¹Sports Engineering Research Centre (SERC), Universiti Malaysia Perlis (UniMAP), Kangar 02600, Malaysia

²Faculty of Electrical Engineering & Technology, Universiti Malaysia Perlis (UniMAP), Kangar 02600, Malaysia

³Faculty of Mechanical Engineering & Technology, Universiti Malaysia Perlis (UniMAP), Kangar 02600, Malaysia

⁴Mechanical Engineering Department, College of Engineering, King Saud University, Riyadh 11421, Saudi Arabia

⁵King Salman Center for Disability Research, Riyadh 11614, Saudi Arabia

⁶Department of Mechanical Engineering, Faculty of Science and Technology, Tokyo University of Science, Chiba, Japan

Correspondence to:

Muhammad Farzik Ijaz*, e-mail: mijaz@ksu.edu.sa, Tel.: +966 11 467 6650, Fax: +966 11 467 6652

Received: April 30 2023; Revised: August 3 2023; Accepted: August 5 2023; Published Online: August 25 2023

ABSTRACT

Sprained ankles are the most commonly diagnosed injury seen by healthcare providers and are projected to account for up to 30% of sports medicine injuries, with lateral ankle sprain being the most common type. Ankle injuries necessarily involve motion assistance to regain mobility, but physiotherapists are typically able to provide rehabilitation only for one patient at each session. Numerous robotic rehabilitation strategies have been proposed in recent years; however, most of the designs have some limitations such as requiring the patient to sit or stand still. Hence, this study aims to develop a conceptual design and simulation of a compact wearable robot in aiding ankle motion for rehabilitation and training purposes. The cable-driven parallel architecture used in the construction of the cable-driven ankle rehabilitation robot allows for the exercise of the human ankle's range of motion (ROM) to be maximized. The morphological chart analysis was created to explore the possible solutions to the design development for the ankle rehabilitation device, and the final design was decided using the Pugh method. A three-dimensional model of the proposed design was visualized in SolidWorks to analyze the inverse kinematics, trajectory simulation and cable length analysis. The feasibility of the ankle rehabilitation robot was examined against the simulation and was found to meet the requirements for performing effective ROM exercises for ankle rehabilitation. The proposed design could potentially be used for passive ankle joint rehabilitation exercise in specific ROM, particularly for bedridden patients.

KEYWORDS

bedridden, cable-driven robot, ankle rehabilitation robot, biomechanics

INTRODUCTION

The ankle joint can be considered an essential part of human movement as it supports the entire body's weight (Li et al., 2021). Because of this, it is also frequently subjected to common injuries such as sprains and fractures. These injuries greatly affect the patient's mobility. Rehabilitation and motion assistance at the early stage are necessary to help regain full mobility (Chen et al., 2019). Physiotherapist usually performs the rehabilitation movement in one-on-one sessions because full attention from the physiotherapist is required. Passive motion for patients with injured ankles during ankle rehabilitation exercises includes adduction–abduction, dorsiflexion–plantarflexion and inversion–eversion motions. Thus, to enable a physiotherapist to

attempt multiple patients simultaneously, a robot-aid rehabilitation device is an efficient alternative. This device could potentially help the physiotherapist teleoperate the robots from a remote location.

Several research works have reported the feasibility of the robot for assisting patients in recovering their ankle mobility (Shi et al., 2019; Kian et al., 2022). However, most of these robots for ankle rehabilitation are based on static platform designs. These solutions require patients to be capable of standing up and are featured with devices that are large and immobile (Jamwal et al., 2016; Chang and Zhang, 2019; Zhang et al., 2019). To overcome this limitation, ankle exoskeletons and orthoses are examples of wearable

designs that have been proposed by several research works (Cafolla et al., 2018; Cafolla et al., 2019; Zuccon et al., 2020). These wearable rehabilitation systems are heavy and stiff, though, and still need support from the sides to support their weight. Due to their lightweight construction, cable-driven robots might offer a solution to this problem (Cafolla et al., 2018). Cable-driven mechanisms provide a higher payload-to-weight ratio, and the design can be made lighter compared to rigid-based robots (Laribi and Ceccarelli, 2021). They have brought forward static (non-portable) designs for ankle rehabilitation (Shahrol et al., 2018; Laribi and Ceccarelli, 2021).

A restrained platform is described in the following report (Von Teh et al., 2020). However, due to the proposed device's attachment to the patient's foot rather than the shank, its four cables are insufficient to permit complete control of the ankle joint. Hence, the present study was undertaken to develop a conceptual design of a cable-driven ankle rehabilitation robot for bedridden patients that is capable of total control of the ankle joint. Before designing the conceptual robot, the specific requirements and problems of the cable-driven ankle rehabilitation robot must be identified and solved for this work. By focusing on the biomechanics of the human ankle joint's mobility, the motion requirements were established.

The aim of this study is to develop a novel conceptual design for a cable-driven ankle rehabilitation robot specifically tailored for bedridden patients. The proposed robot must overcome the limitations of existing static platform designs and wearable systems by providing total control of the ankle joint while remaining lightweight and portable. To achieve this aim, the specific requirements and challenges of the cable-driven ankle rehabilitation robot must be identified and addressed. By focusing on the biomechanics of the human ankle joint's mobility, the motion requirements for effective ankle rehabilitation will be established.

METHODOLOGY

Human ankle motion requirements

The range of motion (ROM) for assistance in smaller ranges is considered mainly for rehabilitation applications as well as assisting the elderly. The movements the ankle must accomplish are considered in the requirements for design and operation. The ankle would move in the desired manner entirely passively.

In unloaded ankle motion (in air), the action size can range from 1 to 20 N and can be scaled up to 100 N if the user is putting load on the foot. The usage of this device is planned for bedridden patients where the ankle is unloaded (no pressure is put on the feet). With exercise training being done to strengthen the ankle muscular system, it can be assumed that the force required for the device's development is <30 N.

According to Saglia et al. (2013), only two rotational degrees of freedom (DOFs) are needed for exercises of the ankle. However, it can be observed that some ankle rehabilitation robots utilize 3 DOFs and 4 DOFs in their design.

Design criteria for concept generation

The main desired criterion for this device is to focus on ankle rehabilitation devices that can be worn and operated on a bedridden patient; therefore, mobility is not the main concern. However, the priority criteria for the conceptual design of this device are safety, portability, ease to use and treatment efficacy.

In figuring out the configuration of the device, consideration has been made to the mechanism design, actuation system and user interface (UI). Since it is desired for the ankle rehabilitation robot to be portable, the pneumatic system might not fit the criteria. Pneumatic actuation requires a compressor, which increases the number of equipment needed for the device to operate. Using an electrical actuator remedies the issues of portability.

The assessment of the desired criteria of the bedridden ankle rehabilitation device was made in comparison to the current existing robot-aided ankle rehabilitation devices. These devices are the Wearable Ankle Rehabilitation Robotic Device, the Optimal Design of Cable Driven Ankle Rehabilitation Robot (ARBOT) and the Optimal Design of Cable Driven Ankle Rehabilitation Robot (Saglia et al., 2013; Von Teh et al., 2020). The description of assessment criteria is available in Appendix 1.

Morphological chart

The criteria required for designing the ankle rehabilitation robot generated by the morphological chart are listed in Table 1. Five concept designs were generated by mixing and matching the criteria.

Selection criteria










Table 2 shows the selection criteria for the ideal concept design that was generated. These criteria were considered as a basis for the overall ankle rehabilitation robot's usability.

Pugh chart concept selection

The Pugh method was utilized for the selection of the conceptual design of the ankle rehabilitation device. This method is typically employed to decide on a design at the early conceptual design stage. It is a graphical method that employs a matrix with columns for the concept and rows for the decision criteria. With this approach, the best concept design was chosen.

The reference design (Datum) is a 3-DOF ankle rehabilitation robot developed by Jamwal et al. (2016) and was compared with another concept design. Each concept design was compared to the Datum. The score will be zero if the concept design's criteria match those of the Datum. The design will receive a +1 if it is superior to the Datum. A score of -1 is given when the concept design is inferior to the Datum. Table 3 lists the concept design selection based on the Pugh method, while Table 4 lists the scale for concept design selection.

Table 1: Morphological chart.

Criteria	Parameters				
Brace type					
Mechanism	Flipflop/sandal Screw mechanism	Foot plate Helical gear Pneumatic	Heel brace Belts Cams	Full brace Lever mechanism Peaucellier cell	Toe-socks/compress Cables PAM (pneumatic artificial muscle) Pulleys
Actuation point	2-way cylinder Spring mechanism Flex muscle (front)	Shafts and gears Hydraulic Flex muscle (back)	Internal gear Swing arm quick return Relax muscle	Spur gears Rack and pinion Horizontal muscle (on ankle)	Flex and relax (front and back)
Actuator	DC rotary motor	Solenoids	Hydraulic cylinder	Shape memory alloy	SEA (series elastic actuator)
Torque application	PMA (pneumatic muscle actuator) 				
Interface	Tension near toes PC	Midfoot tension Control panel (wire connection)	Pulleys	Gear near ankle Remote control (wireless transmission)	Attached muscle
Sensors	Force sensor	Optical encoder	Pressure sensor	Linear potentiometer	Displacement sensor
	Torque sensor	Magnetic rotary encoder	Axis inclinometers	Proximity sensor	Accelerometer
DOF	1	2	3	4	4
Materials	Plastic	Aluminium	Nylon	Neoprene Fibreglass	Spandex

Abbreviation: DOF, degrees of freedom.

Table 2: Selection criteria.

Criteria item	Description
Easy to build	Ease of technology and implementation
Cost effective	Applicable for both production and prototyping
Ease of application	Easy for user to apply and with one person
Comfortable	Pressure, heat, weight, etc.
User friendly	Interface eases uses and easy to adjust
Effective	Adequate performance to accomplish proper ankle treatment exercises and motion
Reconfigurable	Can adjust manually or automatically
Safe	Safe for operation and build
Portability	Ease of use at different location with little adjustment

Inverse kinematics

The proposed cable-driven robot’s inverse kinematics are comparatively straightforward and offer a singular cable length solution for a specific end effector pose. The moving platform (MP) pose shown in Figure 1 was used to calculate the wire lengths used in the discussion that follows.

A spherical joint with limited angular motion can estimate the kinematic behavior of the ankle motion. Using an S-4SPS architecture, the proposed solution can be modeled as a lower-mobility parallel mechanism with 3-DOF. The connection points between the MP and fixed platform (FP) are denoted

by a_i^s and b_i^f , respectively. The FP’s connection points are all in the same plane ($Z_o = 0$) and are located at a radial distance r_b from the coordinate system at O . Position vectors (b_i^o) of point b_i^f s on the FP are defined by equation (1):

$$b_i^o = \begin{bmatrix} r_b \cos \beta_i \\ r_b \sin \beta_i \\ 0 \end{bmatrix} \quad (1)$$

where the radial distance from the origin, O , is indicated by r_b . The variable β_i displays the angle between the respective axes at which point b_i on FP is located. The connecting

Table 3: Pugh method for concept design selection.

Function	Datum	Concept design				
		A	B	C	D	E
Secure foot	Steel platform	Shank brace with foot plate	Full brace	Toe-socks/ compress	Heel brace	Shank brace with foot plate
Mechanism	Pulleys	Bowden cable	Spring mechanism	Pulleys	2-way cylinder	Air muscle
Actuator	DC motor	DC motor	DC motor	SEA	PMA	Soft pneumatic actuator
Actuate muscle	Flex and relax (front and back)	Flex and relax (front and back)	Flex muscle (front)	Horizontal muscle (on ankle)	Flex muscle (back)	Flex muscle (back)
Torque application	Platform under foot	Pulleys	Midfoot tension	Attached muscle	Gear near ankle	Tension near heel
Interface System	PC	PC	Sound	PC	Remote control	Control panel
Sense motion data	Force sensor and accelerometer	Accelerometer and IMU	Displacement and torque sensor	Optical encoder and pressure sensor	Proximity sensor	Accelerometer and force sensor
DOF	4	3	3	3	2	2
Material	Steel	Nylon and plastics	Steel and nylon	Neoprene, spandex, and fiberglass	Nylon	Neoprene and fibreglass.

Abbreviations: DOF, degrees of freedom; PMA, pneumatic muscle actuator; SEA, series elastic actuator.

Table 4: Pugh scale.

Scale	Description
+2	Much better than datum
+1	Better than datum
0	Same as datum
-1	Worse than datum
-2	Much worse than datum

points for the MP are similarly situated on a circle of radius r_a , and the coordinate frame attached to O_e is roughly 495 mm above the MP's center of mass (M_c) (with reference to the position of ankle joint which is about 495 mm above the moving plate level).

Equation (2) can be used to determine the position vectors (a_i^e) of the four connection points on the MP:

$$a_i^e = \begin{bmatrix} r_a \cos \alpha_i \\ r_a \sin \alpha_i \\ -495 \end{bmatrix} \quad (2)$$

The variable β_i indicates the angle at which the point a_i on the MP is in relation to each of its axes. As shown in equation (3), a system of four equations can be used to express the cable length position vectors in terms of end poses:

$$L_i^o = P_e^o + R_e^o a_i^e - b_i^o \quad i = 1, \dots, 4 \quad (3)$$

where P_e^o is the position vector of point O_e relative to O . R_e^o is the rotational transformation matrix of MP with respect to FP using a fixed axis rotation sequence of and about the ψ , θ and ϕ and X_o , Y_o and Z_o axes, respectively, and can be written as in equation (4):

$$R_e^o = \begin{bmatrix} \cos \phi \cos \theta & -\sin \phi \cos \psi + \cos \phi \sin \theta \sin \psi & \sin \phi \sin \psi + \cos \phi \sin \theta \cos \psi \\ \sin \phi \cos \theta & \cos \theta \cos \psi + \sin \phi \sin \theta \sin \psi & -\cos \phi \sin \psi + \sin \phi \sin \theta \cos \psi \\ -\sin \theta & \cos \theta \sin \psi & \cos \theta \cos \psi \end{bmatrix} \quad (4)$$




Figure 1: A sketch of cable and position vectors of connection points on the fixed shank platform and moving foot platform (Jamwal et al., 2016).

RESULTS

The results obtained in this study can be discussed in four parts: (1) the generated concept from the morphological chart

Table 5: Comparison between the generated conceptual design.

Design	Dimension (cm)	DOF	Actuator	Mechanism	Interface	Sensor	Working mode
A	50 × 10 × 25	3	Stepper motor	Pulleys (Bowden cable)	PC	IMU	Lying/standing/sitting
B	31 × 26 × 40	3	DC motor	Spring	Sound	Displacement	Sitting/standing
C	27 × 14 × 32	3	SEA	Pulleys	PC	Optical and torque	Lying/standing/sitting
D	37 × 20 × 35	2	Pneumatic	Pneumatic cylinder	PC	Proximity	Sitting/standing
E	34 × 15 × 25	2	Soft pneumatic actuator	Air muscles	Control panel	Accelerometer and force	Lying/standing/sitting

Abbreviations: DOF, degrees of freedom; SEA, series elastic actuator.

and Pugh method; (2) the selected design and the mechanism generated from the Pugh method; (3) the trajectory simulation of the chosen design; and (4) the performance analysis of the generated concept based on inverse kinematics.

Generation of conceptual design

Table 5 displays the five conceptual designs for ankle rehabilitation robots that have been created and visualized using SolidWorks based on the criteria of the morphological chart.

An ankle rehabilitation robot with a 3-DOF cable-driven parallel mechanism is Design A. The design measures 50 cm by 10 cm by 25 cm. On the sides of the foot plate, close to the ankle and toes, four Bowden cables with stepper motor mechanisms are attached to the shank strap. The motors will pull the Bowden cable to manipulate the moving foot platform. The design has an accelerometer and IMU sensors with a PC as the UI. This design uses nylon plastics, a soft strap for the adjustable part of the shank and a more solid piece for the foot plate. This design can perform ankle rehabilitation exercise motions like inversion, eversion, dorsiflexion, plantarflexion, adduction and abduction and is suitable for bedridden use.

Design B is a 3-DOF full-brace robot-aided ankle rehabilitation device. The dimensions are 31 cm × 26 cm × 40 cm. This device uses two spring mechanisms placed at the front and back of the foot plate to allow movement for dorsiflexion and plantarflexion. This robot uses a DC motor to actuate the springs. For ankle rehabilitation exercises in the sitting and standing modes, this design is better suited. Only plantarflexion and dorsiflexion ankle exercise motions are doable by this design. The ankle robot uses sound as the UI and has displacement sensors placed to detect the distance of placement. Design B is made of steel frame and nylon for padding.

Design C is a 3-DOF ankle rehabilitation robot. The dimensions of Design C are 27 cm × 14 cm × 32 cm. It has a toe-socks foot sleeve design. This design uses a series elastic actuator (SEA) for actuation of pulleys. Three pulleys are placed near the dorsum of the foot to replicate the movement of the dorsal muscles. This design allows for the practice of ankle rehabilitation exercise motions like inversion, eversion, dorsiflexion and plantarflexion. For gait sensing, optical encoder and torque sensors are used to detect and interpret robot movement. The PC is the UI for this robot. The robot in this design uses lightweight, sturdy materials such as neoprene, spandex and fiberglass for the frame. This

design is good for bedridden patients’ use as it can function in any position.

Design D is a 2-DOF heel brace ankle rehabilitation robot with dimensions 37 cm × 20 cm × 35 cm. This device uses a two-way pneumatic cylinder mechanism. This device can perform plantar flexion and dorsiflexion ankle rehabilitation exercises. It can be used in sitting and standing positions, though using this design while lying down might not be feasible due to the pneumatic requirements. This ankle robot uses nylon for the brace padding and plastic for the frame. It has a proximity sensor to detect gait.

Design E is a 2-DOF over-actuated parallel mechanism ankle rehabilitation robot with dimensions 34 cm × 15 cm × 25 cm. It is a shank brace with a foot plate design. The robot uses air muscles affixed to the sides and back of the brace actuated by a soft pneumatic actuator. The robot can perform dorsiflexion/plantarflexion and eversion/inversion ankle motions. This device can be worn for standing and bedridden uses. It uses accelerometer and force sensor for motion detection and a control panel as the UI. The materials used for this device are neoprene and fiberglass.

From the generated selection criteria, the weights of each criterion were given, and points attributed to each criterion were generated as the table against the Datum. Based on the generated designs, the selection criteria are compared, and points are given. Designs with better than the Datum get a positive point, and the magnitude of the point dictates how good the design is. From Table 6, it can be observed that Design A has criteria that are the best. The design has a simple build, and that also means that it is cost effective as well. It did not score well for comfort, but its simplistic build meant that it is user-friendly. Criteria that are the same as the Datum are scored at zero (0).

Table 6: Selection criteria for the chosen design.

Selection criteria	Datum	A	B	C	D	E
Easy to build	3-DOF	+2	+1	-1	+1	+1
Cost effective	ankle	+2	0	-2	-1	+1
Ease of application	robot	+2	+1	-1	+1	+1
Comfortable		+1	+1	+2	+1	0
User friendly		+2	+1	+1	+1	+1
Effective		0	0	0	0	0
Reconfigurable		+1	-1	-1	+1	+1
Safe		+2	+1	+2	+2	+2
Portability		+2	+1	+2	+2	+2

Abbreviation: DOF, degrees of freedom.

Selected design

The ankle robot is designed to give the ankle joint 3 DOFs in rotation. The apparatus has two parallel platforms: one FP affixed to the shank and a moving foot platform where the foot is placed. The leg support structure is ring-shaped or straps that is fixed to patient's leg. The cable lengths are varied by the motors. The overview revealing the conceptual layout of the wearable ankle rehabilitation device with main components and corresponding 3D model of the proposed conceptual design are presented in Figures 2 and 3 respectively.

A mechanical structure with Bowden cables and servo-motor will guide the relative motion of the cables. For the operational control of the device, sensors can be used. These sensors may also provide monitoring for ankle medical issues, such as muscle reaction, blood pressure and stiffness measurement. Sensors can also be installed on the device itself or worn by the patient to provide information for the control system. These sensors, which enable gait analysis, include accelerometers and IMUs, as reported in Jamwal et al. (2016).

Generation of MATLAB model

MATLAB model for the proposed design of the ankle robot is shown in Figure 4. This model is developed based on the inverse kinematic analysis that has been performed. The actuation cables are positioned at the FP in Figure 5 as a base and attached to the MP to allow movement for the affected foot. The actuation cable used in this mechanism is proposed to be Bowden cable mechanism, where the FP attached to the shank will serve as the base of the moving Bowden cable, and an external motor will perform the pulling motion to permit motion on the MP.

The fixed points are placed at the blue dots. This MATLAB model is made according to the dimensions of Concept Design A. Therefore, equation (2) was used to determine the new vector position for the MP after motions of the ankle have been performed.




Figure 2: Proposed design: Conceptual layout of the wearable ankle rehabilitation device with the main components.




Figure 3: 3D model of the proposed conceptual design.

Trajectory simulation

The trajectory simulation for the MATLAB model of the ankle rehabilitation robot is first set at the origin position, as shown in Figure 6. This origin position is at $(-250, -100, 0)$ for a^1 ; $(100, -100, 0)$ for a^2 ; $(-250, 100, 0)$ for a^3 ; and $(100, 100, 0)$ for a^4 . The trajectory simulation presented with this MATLAB simulation is based on the 3D SolidWorks design proposed and the inverse kinematics derived from the design. The values for the angles used are based on the required ROM exercises that are effective for ankle rehabilitation. These ranges are simulated from Figures 7 to 10 for the inversion, eversion, dorsiflexion, plantarflexion and abduction motions. The initial position for the MP is represented by the black rectangular, while the new position after angle values are placed is represented by the blue rectangular. The green dotted lines represent the cables attached to the two platforms. The red rectangle represents the FP that bounds the motion of the cables and MP.

Cable length analysis

Table 7 shows the cable length for all the axes of the ROM consisting of the inversion/eversion on the x -axis, dorsiflexion/plantarflexion on the y -axis and adduction/abduction on the z -axis. The result shows that the longest cable extension at any ROM for the ankle robot is at plantarflexion with 411.1228 mm extension on the -45 degree angle for both L1 and L3. Meanwhile, the shortest




Figure 4: MATLAB model of the ankle rehabilitation robot.




Figure 5: Dimensions for FP and MP. Abbreviations: FP, fixed platform; MP, moving platform.




Figure 6: Simulation for the inversion range on the x-axis.

cable retraction is in the same range for L2 and L4 at 183.0615 mm.

DISCUSSION AND CONCLUSION

In this study, Design A was selected from five other generated designs using the morphological chart where the features required and needed were picked from an array of possible selection of features. Using the Pugh method to select the best design generated from the morphological chart, careful consideration was given in the selection process.

The finalized design consists of a parallel cable-driven robot with an FP attached to the leg's shank acting as the FP, and an MP, or end effector, located at the foot. This design uses Bowden cables for actuation where the external actuation is done with a DC motor. This design is suitable for




Figure 7: Simulation for the eversion range on the x-axis. (a) 14.5 degrees and (b) 22 degrees.




Figure 8: Simulation for the dorsiflexion range on the y-axis. (a) -10 degrees and (b) -17 degrees.




Figure 9: Simulation for the plantarflexion range on the y-axis. (a) 20 degrees and (b) 30 degrees.

different positions whether for bedridden use, standing or sitting. With the actuator not being attached directly to the orthosis itself, the system allows for more comfort as there is

no cumbersome mechanism attached to the ankle robot itself. The design uses sliders as a proponent for the structural integrity of the ankle rehabilitation robot. From the generated 3D




Figure 10: Simulation for the abduction range on the z-axis. (a) -37 degrees, (b) -45 degrees, (c) 22 degrees and (d) 35 degrees.

Table 7: Cable length for all axes maximum and minimum.

Axes	Range of motion	Angle (degrees)	(mm)			
			L1	L2	L3	L4
0	Origin	0	307.5759	307.2163	307.2163	307.5759
x	Inversion	14.5	345.2694	279.3925	267.4697	230.2655
		22	363.3602	291.1930	246.3893	217.5609
	Eversion	-10	280.0725	238.0053	333.9195	272.0502
		-17	260.4416	225.9944	351.4196	283.3913
y	Dorsiflexion	20	252.0047	288.1054	252.0047	288.1054
		30	224.1741	303.0548	224.1741	303.0548
	Plantarflexion	-37	393.9524	193.9330	395.9524	193.9330
		-45	411.1228	183.0615	411.1228	183.0615
z	Adduction	22	313.2657	258.8257	313.267	258.8257
		35	321.5218	264.0897	321.5218	264.0897
	Abduction	-15	310.2514	256.9116	310.2514	256.9116
		-25	314.8780	259.8512	314.8780	259.8512

design using SolidWorks, the kinematic model, the inverse kinematic model was then developed to have the mathematical models plugged into MATLAB, where a simulation was then programmed using the plot3 functions to represent the platforms and projected cable behavior of the robot.

In this simulation, the three axes of ankle motion—inversion/eversion, dorsiflexion/plantarflexion and abduction/

adduction—were used to represent the lengthening of the cable and test the viability of a robot design for achieving the required ROM exercises. From the MATLAB simulation and the data derived from the program, it can be observed that Concept Design A could achieve the full ROM for effective ankle rehabilitation exercises. In conclusion, this ankle robot rehabilitation design is successful in fulfilling

the design requirements for an effective ankle rehabilitation robot.

FUNDING

The authors extend their appreciation to the King Salman Center For Disability Research for funding this work through Research Group no KSRG-2022-033.

AUTHOR CONTRIBUTIONS

NH, MNS and SNB conceptualized the study. NH, KSB and MFI performed the methodology. NH investigated the study. HT and MNS were responsible for the resources. NH prepared

the original draft (writing). HT, KSB and MFI reviewed and edited the manuscript. NH and MFI were in-charge of visualization. MFI performed funding acquisition.

CONFLICTS OF INTEREST

The authors declare no conflicts of interest in association with the present study.

ACKNOWLEDGMENTS

The authors extend their appreciation to the King Salman Center for Disability Research for funding this work through Research Group no KSRG-2022-033.

REFERENCES

- Cafolla D., Russo M. and Carbone G. (2018). Design and validation of an inherently safe cable-driven assisting device. *Int. J. Mech. Control*, 19(1), 23-32.
- Cafolla D., Russo M. and Carbone G. (2019). CUBE, a cable-driven device for limb rehabilitation. *J. Bionic Eng.*, 16(3), 492-502. 10.1007/s42235-019-0040-5.
- Chang T.C. and Zhang X.D. (2019). Kinematics and reliable analysis of decoupled parallel mechanism for ankle rehabilitation. *Microelectron. Reliab.*, 99, 203-212. 10.1016/j.microrel.2019.05.016.
- Chen E.T., Borg-Stein J. and McInnis K.C. (2019). Ankle sprains: evaluation, rehabilitation, and prevention. *Curr. Sports Med. Rep.*, 18(6), 217-223. 10.1249/JSR.0000000000000603.
- Jamwal P.K., Hussain S., Ghayesh M.H. and Rogozina S.V. (2016). Impedance control of an intrinsically compliant parallel ankle rehabilitation robot. *IEEE Trans. Ind. Electron.*, 63(6), 3638-3647. 10.1109/TIE.2016.2521600.
- Kian A., Widanapathirana G., Joseph A.M., Lai D.T.H. and Begg R. (2022). Application of wearable sensors in actuation and control of powered ankle exoskeletons: a comprehensive review. *Sensors*, 22(6), 2244. 10.3390/s22062244.
- Laribi M.A. and Ceccarelli M. (2021). Design and experimental characterization of a cable-driven elbow assisting device. *J. Med. Devices*, 15(1), 014503. 10.1115/1.4049529.
- Li H., Xia C., Li F., Yin J., Wang X., Cheng J., et al. (2021). Motion simulation of bionic auxiliary device for ankle rehabilitation based on small driving moment function. *IOP Conf. Ser.: Earth Environ. Sci.*, 1802(4), 042086. 10.1088/1742-6596/1802/4/042086.
- Ren Y., Wu Y.N., Yang C.Y., Xu T., Harvey R.L. and Zhang L.Q. (2017). Developing a wearable ankle rehabilitation robotic device for in-bed acute stroke rehabilitation. *IEEE Trans. Neural Syst. Rehabil. Eng.*, 25(6), 1-1. 10.1109/TNSRE.2016.2584003.
- Saglia J.A., Tsagarakis N.G., Dai J.S. and Caldwell D.G. (2013). Control strategies for patient-assisted training using the ankle rehabilitation robot (ARBOT). *IEEE/ASME Trans. Mechatronics*, 18(6), 1799-1808. 10.1109/TMECH.2012.2214228.
- Shahrol M.N., Basah S.N., Basaruddin K.S., Ahmad W.K.W. and Ahmad S.A. (2018). Modelling of a cable-driven ankle rehabilitation robot. *J. Telecommun. Electron. Comput. Eng.*, 10(1-10), 53-59.
- Shi B., Chen X., Yue Z., Yin S., Weng Q., Zhang X., et al. (2019). Wearable ankle robots in post-stroke rehabilitation of gait: a systematic review. *Front. Neurobot.*, 13, 63. 10.3389/fnbot.2019.00063.
- Von the C.Y., Bin Shahrol Aman M.N.S., Mustafa W.A. and Bin Ahmad S.A. (2020). Conceptual design for ankle rehabilitation robot by using morphological chart and Pugh method. *IOP Conf. Ser.: Mater. Sci. Eng.*, 932(1), 012062. 10.1088/1757-899X/932/1/012062.
- Zhang M., Mcdaid A., Veale A.J., Peng Y. and Xie S.Q. (2019). Adaptive trajectory tracking control of a parallel ankle rehabilitation robot with joint-space force distribution. *IEEE Access*, 7, 85812-85820. 10.1109/ACCESS.2019.2925182.
- Zuccon G., Bottin M., Ceccarelli M. and Rosati G. (2020). Design and performance of an elbow assisting mechanism. *Machines*, 8(4), 68. 10.3390/machines8040068.

APPENDIX 1

DESCRIPTION OF ASSESSMENT CRITERIA

1. **Brace type:** This criterion was used as a point of reference for possible brace design. The type of brace used for the ankle rehabilitation robot for bedridden users is important to consider for safety, comfort and how the mechanism can be affixed to the setup.
2. **Actuation point:** This criterion was considered for the placement of actuation. Determining the placement will ensure the effectiveness of the design in performing the three motion modes of the ankle, wherever possible. Saglia et al. (2013) and Ren et al. (2017) placed the actuators of the device at the side, while Jamwal et al. (2016) placed four actuators along the fixed platform (FP).
3. **Actuation:** This criterion was chosen to take the actuator for the ankle rehabilitation robot into account. For the device developed by Jamwal et al. (2016), air muscles were used for actuation, while Ren et al. apply a planetary gearhead powered by a brushless motor.
4. **Mechanism:** The transmission power of the actuator to the FP was calculated using this criterion. Air muscles were used by Jamwal et al. (2016); their actuating end was connected to the mobile platform via cables, and they were attached to a leg support. Ren et al. used a brushless motor with planetary gearhead and a bevel gear set. The potential mechanisms for the design are cables (pulleys), spurs, two-way cylinder, one-way air muscle, spring-loaded mechanism and gears and shafts. For the mechanism, a combination of two or more types of mechanisms can be chosen.
5. **Sensors:** This criterion was considered when choosing the kind of sensors to be used in the proposed ankle rehabilitation robot to control movement and maintain safety. To quantify motor-driving torque, and joint resistance, Ren et al. placed a force sensor with a 100-lb potential between both the foot bearing adapter and the driving linkage frame (Ren et al., 2017).
6. **User interface:** This type of criterion was used to determine the type of interface for the ankle rehabilitation robot. Jamwal et al. (2016), Ren et al. (2017) and Saglia et al. (2013) selected the PC as the interface. Ren et al. (2017) employed a game format in which, through the preset within the games, the robotic device senses the lack of patient movement through disparity detection between the real ankle position and the target position.
7. **Material:** The robot for ankle rehabilitation was made from the best material according to this criterion. The ankle rehabilitation robot could be made out of plastics, aluminum, nylon, neoprene, spandex and fiberglass.
8. **Degree of freedom (DOF):** This criterion is to observe the number of possible DOFs with the design.